

CFR Code of Federal Regulations Title 21

•  ¹

•  ³ [FDA Home](#) ⁴ [Medical Devices](#) ⁵ [Databases](#)



⁶ [510\(k\)](#) ⁷ [Registration & Listing](#) ⁸ [Adverse Events](#) ⁹ [Recalls](#) ¹⁰ [PMA](#) ¹¹ [Classification](#) ¹² [Standards](#) ¹³
[CFR Title 21](#) ¹⁴ [Radiation-Emitting Products](#) ¹⁵ [X-Ray Assembler](#) ¹⁶ [Medsun Reports](#) ¹⁷ [CLIA](#) ¹⁸ [TPLC](#) ¹⁹

[New Search](#)

[Code of Federal Regulations]
 [Title 21, Volume 8]
 [Revised as of April 1, 2012]
 [CITE: 21CFR882.1870]

[Help](#) ²⁰ | [More About 21CFR](#) ²¹



TITLE 21--FOOD AND DRUGS
 CHAPTER I--FOOD AND DRUG ADMINISTRATION
 DEPARTMENT OF HEALTH AND HUMAN SERVICES
 SUBCHAPTER H--MEDICAL DEVICES

PART 882 -- NEUROLOGICAL DEVICES

Subpart B--Neurological Diagnostic Devices

Sec. 882.1870 Evoked response electrical stimulator.

(a)*Identification.* An evoked response electrical stimulator is a device used to apply an electrical stimulus to a patient by means of skin electrodes for the purpose of measuring the evoked response.

(b)*Classification.* Class II (performance standards).